



**Sex Hormone Binding Globulin (SHBG) Test System**  
**Product Code: 9125-300**

**1.0 INTRODUCTION**

**Intended Use: The Quantitative Determination of Sex Hormone Binding Globulin (SHBG) Concentration in Human Serum by a Microplate Enzyme Immunoassay, Colorimetric.**

**2.0 SUMMARY AND EXPLANATION OF THE TEST**

In order for the sex-steroids testosterone, 5 $\alpha$ -dihydrotestosterone, and 17 $\beta$ -estradiol to reach the tissues, a transport molecule known as the Sex Hormone Binding Globulin (SHBG) is used. No more than 10% of these steroids are actually in plasma unbound and, therefore, biologically active. This protein's steroid binding capacity is highly dependent on the temperature and pH of its environment.<sup>1,2,3</sup> Originating in the liver, this glycoprotein is a 93.4 kDa homodimer. Overall, SHBG is responsible for the balance of steroid levels and in certain instances has an influence on the effects these steroids have on their targets. The serum levels of this globulin have been shown to vary drastically between individuals, even if considered to be in a "normal" state of health.<sup>2,3,4</sup>

Physiological status changes due to hormonal, metabolic, and nutritional factors are reflected in the concentration of SHBG in serum. Rises in oestrogens correlate with an increase in SHBG, whereas a rise in androgens inhibits SHBG production. Aging and conditions like polycystic ovarian syndrome demonstrate this effect.<sup>2,3,5</sup> SHBG levels also correlates strongly with conditions like hyperthyroidism, insulin resistance, central adiposity, and dyslipidemia. More importantly, low levels of SHBG shows an increased risk for diabetes and cardiovascular disease.<sup>5,6,7</sup>

SHBG is also used to calculate the Free Androgen Index (FAI). The calculation method has been applied frequently to the determination of Free Testosterone (FT) levels. In the past, analog methods and dialysis have been used to obtain free testosterone levels, but each has a very striking downfall: analog methods tend to give values substantially lower than the actual clinical status and using dialysis is a laborious technique. This calculated method uses SHBG, total testosterone (T) and albumin concentrations to calculate the free Testosterone in human serum.<sup>8,9</sup>


**3.0 PRINCIPLE**

**Immunoenzymometric assay (TYPE 3):**

The essential reagents required for an immunoenzymometric assay include high affinity and specificity antibodies (enzyme conjugated and immobilized), with different and distinct epitope recognition, in excess, and native antigen. In this procedure, the immobilization takes place during the assay at the surface of a microplate well through the interaction of streptavidin coated on

the well and exogenously added biotinylated monoclonal anti-SHBG antibody.

Upon mixing monoclonal biotinylated antibody, the enzyme-labeled antibody and a serum containing the native antigen, reaction results between the native antigen and the antibodies, without competition or steric hindrance, to form a soluble sandwich complex. The interaction is illustrated by the following equation:



$\text{B}^{\text{m}}\text{Ab}_{(m)}$  = Biotinylated Monoclonal Antibody (Excess Quantity)

$\text{Ag}_{\text{SHBG}}$  = Native Antigen (Variable Quantity)


$\text{EnzAb}_{(m)}$  = Enzyme-Monoclonal Antibody (Excess Quantity)

$\text{EnzAb}_{(m)}\text{-Ag}_{\text{SHBG}}\text{-B}^{\text{m}}\text{Ab}_{(m)}$  = Antigen-Antibodies Sandwich Complex

$k_a$  = Rate Constant of Association

$k_{-a}$  = Rate Constant of Dissociation

Simultaneously, the complex is deposited to the well through the high affinity reaction of streptavidin and biotinylated antibody. This interaction is illustrated below:



$\text{Streptavidin}_{\text{CW}}$  = Streptavidin immobilized on well

Immobilized complex = sandwich complex bound to the well surface

After equilibrium is attained, the antibody-bound fraction is separated from unbound antigen by decantation or aspiration. The enzyme activity in the antibody-bound fraction is directly proportional to the native antigen concentration. By utilizing several different serum references of known antigen values, a dose response curve can be generated from which the antigen concentration of an unknown can be ascertained.

**4.0 REAGENTS**

**Materials Provided:**

**A. SHBG Calibrators – 1ml/vial - Icons A-F**

Six (6) vials of human serum based reference calibrators at concentrations of 0 (A), 10 (B), 25 (C), 50 (D), 100 (E) and 250 (F) nmol/L. A preservative has been added. Store at 2-8°C. (The Calibrators are standardized against WHO's 2<sup>nd</sup> IS 08/266 for SHBG).

**B. SHBG Enzyme Reagent – 12ml/vial - Icon**

One (1) vial of anti-human SHBG-HRP and biotinylated anti-human SHBG presented in a protein-stabilized matrix. A preservative has been added. Store at 2-8°C.

**C. SHBG Diluent – 12ml/vial – Icon**

One (1) vial of a protein-stabilized matrix. A preservative has been added. Store at 2-8°C.

**D. Streptavidin Plate – 96 wells – Icon**

One 96-well microplate coated with streptavidin and packaged in an aluminum bag with a drying agent. Store at 2-8°C.

**E. Wash Solution Concentrate – 20ml/vial - Icon**

One (1) vial containing a surfactant in buffered saline. A preservative has been added. Store at 2-8°C.

**F. Substrate A – 7ml/vial - Icon**

One (1) vial containing tetramethylbenzidine (TMB) in acetate buffer. Store at 2-8°C.

**G. Substrate B – 7ml/vial - Icon**

One (1) vial containing hydrogen peroxide (H<sub>2</sub>O<sub>2</sub>) in acetate buffer. Store at 2-8°C.

**H. Stop Solution – 8ml/vial - Icon**

One (1) vial containing a strong acid (1N HCl). Store at 2-8°C.

**I. Product Instructions.**

**Note 1:** Do not use reagents beyond the kit expiration date.

**Note 2:** Avoid extended exposure to heat and light. **Opened reagents are stable for sixty (60) days when stored at 2-8°C. Kit and component stability are identified on the label.**

**Note 3:** Above reagents are for a single 96-well microplate.

**4.1 Required But Not Provided:**

1. Pipette capable of delivering 0.025 & 0.050ml (25 & 50 $\mu$ l) volumes with a precision of better than 1.5%.
2. Dispenser(s) for repetitive deliveries of 0.100 & 0.300ml (100 & 300 $\mu$ l) volumes with a precision of better than 1.5%.
3. Microplate washers or a squeeze bottle (optional).
4. Microplate Reader with 450nm and 620nm wavelength absorbance capability.
5. Absorbent Paper for blotting the microplate wells.
6. Plastic wrap or microplate cover for incubation steps.
7. Vacuum aspirator (optional) for wash steps.
8. Timer.
9. Quality control materials.

**5.0 PRECAUTIONS**

**For In Vitro Diagnostic Use  
 Not for Internal or External Use in Humans or Animals**

All products that contain human serum have been found to be non-reactive for Hepatitis B Surface Antigen, HIV 1&2 and HCV Antibodies by FDA licensed reagents. Since no known test can offer complete assurance that infectious agents are absent, all human serum products should be handled as potentially hazardous and capable of transmitting disease. Good laboratory procedures for handling blood products can be found in the Center for Disease Control / National Institute of Health, "Biosafety in Microbiological and Biomedical Laboratories," 2nd Edition, 1988, HHS Publication No. (CDC) 88-8395.

**Safe disposal of kit components must be according to local regulatory and statutory requirement.**

**6.0 SPECIMEN COLLECTION AND PREPARATION**

The specimens shall be blood serum and the usual precautions in the collection of venipuncture samples should be observed. For accurate comparison to established normal values, a fasting morning serum sample should be obtained. The blood should be collected in a plain redtop venipuncture tube without additives or anti-coagulants. Allow the blood to clot for samples. Centrifuge the specimen to separate the serum from the cells.

Samples may be refrigerated at 2-8°C for a maximum period of five (5) days. If the specimen(s) cannot be assayed within this time, the sample(s) may be stored at temperatures of -20°C for up to 30 days. Avoid use of contaminated devices. Avoid repetitive freezing and thawing. When assayed in duplicate, 0.050ml (50 $\mu$ l) of the specimen is required.

**7.0 QUALITY CONTROL**

Each laboratory should assay controls at levels in the low, normal and elevated range for monitoring assay performance. These controls should be treated as unknowns and values determined in every test procedure performed. Quality control charts should be maintained to follow the performance of the supplied reagents. Pertinent statistical methods should be employed to ascertain trends. Significant deviation from established performance can indicate unnoticed change in experimental conditions or degradation of kit reagents. Fresh reagents should be used to determine the reason for the variations

**8.0 REAGENT PREPARATION**

1. **Wash Buffer**  
Dilute contents of wash concentrate to 1000ml with distilled or deionized water in a suitable storage container. Store diluted buffer at 2-30°C for up to 60 days.
2. **Working Substrate Solution – Stable** for one (1) year  
Pour the contents of the amber vial labeled Solution 'A' into the clear vial labeled Solution 'B'. Place the yellow cap on the clear vial for easy identification. Mix and label accordingly. Store at 2 - 8°C.
3. **Patient Sample Dilution (1/21)**  
Dispense 0.025ml (25 $\mu$ l) of each patient specimen into 0.500ml (500 $\mu$ l) of serum diluent. Cover and vortex or mix thoroughly by inversion. Store at 2-8°C for up to forty-eight (48) hours.

**Note 1: Do not use the working substrate if it looks blue.**  
**Note 2: Do not use reagents that are contaminated or have bacteria growth.**

**9.0 TEST PROCEDURE**

*Before proceeding with the assay, bring all reagents, reference calibrators and controls to room temperature (20 - 27° C).*

**\*\*Test procedure should be performed by a skilled individual or trained professional\*\***

1. Format the microplates' wells for each serum reference calibrator, control and patient specimen to be assayed in duplicate. **Replace any unused microwell strips back into the aluminum bag, seal and store at 2-8°C.**
2. Pipette 0.025ml (25 $\mu$ l) of the appropriate serum reference calibrator, control or patient sample" into the assigned well. *"Patient sample dilution required – see section 8.0 Reagent Preparation.*
3. Add 0.100ml (100 $\mu$ l) of the SHBG Enzyme Reagent to each well. **It is very important to dispense all reagents close to the bottom of the coated well.**
4. Swirl the microplate gently for 20-30 seconds to mix and cover.
5. Incubate 30 minutes at room temperature.
6. Discard the contents of the microplate by decantation or aspiration. If decanting, tap and blot the plate dry with absorbent paper.
7. Add 0.300ml (300 $\mu$ l) of wash buffer (see Reagent Preparation Section), decant (tap and blot) or aspirate. Repeat two (2) additional times for a total of three (3) washes. **An automatic or manual plate washer can be used. Follow the manufacturer's instruction for proper usage. If a squeeze bottle is employed, fill each well by depressing the container (avoiding air bubbles) to dispense the wash. Decant the wash and repeat two (2) additional times.**
8. Add 0.100ml (100 $\mu$ l) of working substrate solution to all wells (see Reagent Preparation Section). **Always add reagents in the same order to minimize reaction time.**  
**DO NOT SHAKE THE PLATE AFTER SUBSTRATE ADDITION**
9. Incubate at room temperature for fifteen (15) minutes.
10. Add 0.050ml (50 $\mu$ l) of stop solution to each well and gently mix for 15-20 seconds.
11. Read the absorbance in each well at 450nm (using a reference wavelength of 620-630nm to minimize well imperfections) in a microplate reader. **The results should be read within thirty (30) minutes of adding the stop solution.**

**10.0 CALCULATION OF RESULTS**

A dose response curve is used to ascertain the concentration of SHBG in unknown specimens.


1. Record the absorbance obtained from the printout of the microplate reader as outlined in Example 1.
2. Plot the absorbance for each duplicate serum reference versus the corresponding SHBG concentration in nmol/LI on linear graph paper (do not average the duplicates of the serum references before plotting).
3. Draw the best-fit curve through the plotted points.
4. To determine the concentration of SHBG for an unknown, locate the average absorbance of the duplicates for each unknown on the vertical axis of the graph, find the intersecting point on the curve, and read the concentration (in nmol/L) from the horizontal axis of the graph (the duplicates of the unknown may be averaged as indicated). In the following example, the average absorbance (0.698) intersects the dose response curve at 42.06 nmol/L SHBG concentration (See Figure 1).

**Note:** Computer data reduction software designed for ELISA assays may also be used for the data reduction. **If such software is utilized, the validation of the software should be ascertained.**

### EXAMPLE 1

I.D.	Well	Abs	Mean Abs (B)	Conc (nmol/L)
Cal A	A1	0.005	0.006	0
	B1	0.007		
Cal B	C1	0.161	0.160	10
	D1	0.159		
Cal C	E1	0.404	0.407	25
	F1	0.409		
Cal D	G1	0.826	0.826	50
	H1	0.826		
Cal E	A2	1.425	1.433	100
	B2	1.441		
Cal F	C2	2.689	2.674	250
	D2	2.658		
Patient 1	A3	0.717	0.698	42.06
	B3	0.679		

\*The data presented in Example 1 and Figure 1 is for illustration only and **should not** be used in lieu of a standard curve prepared with each assay.



### 11.0 Q.C. PARAMETERS

In order for the assay results to be considered valid the following criteria should be met:

1. The absorbance (OD) of calibrator 'A' should be  $\leq 0.1$ .
2. The absorbance (OD) of calibrator 'F' should be  $\geq 1.3$ .
3. Four out of six quality control pools should be within the established ranges.

### 12.0 RISK ANALYSIS

The MSDS and Risk Analysis Form for this product are available on request from Monobind Inc.

#### 12.1 Assay Performance

1. It is important that the time of reaction in each well is held constant to achieve reproducible results.
2. Pipetting of samples should not extend beyond ten (10) minutes to avoid assay drift.
3. Highly lipemic, hemolyzed or grossly contaminated specimen(s) should not be used.
4. If more than one (1) plate is used, it is recommended to repeat the dose response curve.
5. The addition of substrate solution initiates a kinetic reaction, which is terminated by the addition of the stop solution. Therefore, the substrate and stop solution should be added in the same sequence to eliminate any time-deviation during reaction.
6. Plate readers measure vertically. Do not touch the bottom of the wells.
7. Failure to remove adhering solution adequately in the aspiration or decantation wash step(s) may result in poor replication and spurious results.
8. Use components from the same lot. No intermixing of reagents from different batches.
9. Accurate and precise pipetting, as well as following the exact time and temperature requirements prescribed are essential.

Any deviation from Monobind's IFU may yield inaccurate results.

10. All applicable national standards, regulations and laws, including, but not limited to, good laboratory procedures, must be strictly followed to ensure compliance and proper device usage.
11. It is important to calibrate all the equipment e.g. Pipettes, Readers, Washers and/or the automated instruments used with this device, and to perform routine preventative maintenance.

12. Risk Analysis- as required by CE Mark IVD Directive 98/79/EC - for this and other devices, made by Monobind, can be requested via email from [Monobind@monobind.com](mailto:Monobind@monobind.com).

#### 12.2 Interpretation

1. **Measurements and interpretation of results must be performed by a skilled individual or trained professional.**
2. Laboratory results alone are only one aspect for determining patient care and should not be the sole basis for therapy, particularly if the results conflict with other determinants.
3. The reagents for AccuBind® ELISA procedures have been formulated to eliminate maximal interference; however, potential interaction between rare serum specimens and test reagents can cause erroneous results. Heterophilic antibodies often cause these interactions and have been known to be problems for all kinds of immunoassays (Boscato LM, Stuart MC. "Heterophilic antibodies: a problem for all immunoassays" *Clin.Chem.* 1988:3427-33). For diagnostic purposes, the results from this assay should be used in combination with clinical examination, patient history and all other clinical findings.
4. For valid test results, adequate controls and other parameters must be within the listed ranges and assay requirements.
5. If test kits are altered, such as by mixing parts of different kits, which could produce false test results, or if results are incorrectly interpreted, **Monobind shall have no liability.**
6. If computer controlled data reduction is used to interpret the results of the test, it is imperative that the predicted values for the calibrators fall within 10% of the assigned concentrations.
7. Serum SHBG concentration is dependent upon a multiplicity of factors: including if the patient is sensitized, how many times the patient has been exposed to a specific allergen etc. Total SHBG concentration alone is not sufficient to assess the clinical status. All the clinical findings especially specific allergy testing should be taken into consideration while determining the clinical status of the patient.
8. Since all atopic reactions are not SHBG mediated, all relevant clinical information should be taken into consideration before making any determination for patients who may be in the normal range.

### 13.0 EXPECTED RANGES OF VALUES

In agreement with established reference intervals for a "normal" population the expected ranges for the SHBG AccuBind® ELISA Test System are detailed in Table 1.

POPULATION	RANGE
Males	10 – 57 nmol/L
Females (non-pregnant)	18 – 144 nmol/L

It is important to keep in mind that establishment of a range of values, which can be expected to be found by a given method for a population of "normal" persons, is dependent upon a multiplicity of factors: the specificity of the method, the population tested and the precision of the method in the hands of the analyst. For these reasons, each laboratory should depend upon the range of expected values established by the manufacturer only until an in-house range can be determined by the analysts using the method with a population indigenous to the area in which the laboratory is located.

### 14.0 PERFORMANCE CHARACTERISTICS

#### 14.1 Precision

In order to validate the intra-assay precisions of the SHBG AccuBind® ELISA test system, twenty replicates of each of three pooled sera (low medium and high ranges of the dose response

curve) were assayed in the same assay. An intra-assay precision of 1.5 to 2.6% was obtained.

TABLE 2  
Intra-Assay Precision (in nmol/L)


SAMPLE	N	X	$\sigma$	C.V.%
Control Level 1	24	18.331	0.764	4.2
Control Level 2	24	49.516	2.128	4.3
Control Level 3	24	83.955	4.799	5.7

#### 14.2 Sensitivity

The SHBG AccuBind® ELISA test system has a sensitivity of 0.0122nmol/L. The sensitivity was ascertained by determining the variability of the 0 nmol/L serum calibrator and using the  $2\sigma$  (95% certainty) statistics to calculate the minimum dose.

#### 14.3 Accuracy

The SHBG AccuBind® ELISA test system was compared with a reference method. Biological specimens with SHBG levels in the low, medium and high ranges were used; the values ranged from 4.6 to 184.0 nmol/L. The total number of such specimens was 60. The least square regression equation and the correlation coefficient were computed for this SHBG AccuBind® ELISA method in comparison with the reference method (Table 3).



Parameter	Monobind	Reference
Low	8.86	4.6
High	160.39	184.02
Mean	56.09	49.79
Intercept		1.01
Slope		6.3037
Corr (R <sup>2</sup> )		0.9269

Only slight amounts of bias between this method and the reference method are indicated by the closeness of the mean values. The least square regression equation and correlation coefficient indicates excellent method agreement.

#### 14.4 Specificity

The specificity of the SHBG AccuBind® ELISA test system, to closely related immunoglobulins was evaluated by adding those at twice the physiological concentrations to a serum matrix.

TABLE 4  
SHBG ELISA Specificity and Cross Reactivity

SUBSTANCE	Cross Reactivity
Corticosteroid Binding Globulin	< 0.3
Thyroxine Binding Globulin	< 0.08

### 15.0 REFERENCES

1. Hammond, GL. *J.Steroid Biochem.* **1985**, 23, 451-460.
2. Metzger, J. *Biochemistry*, **2003**, 42, 13735-13745.
3. Anderson, DC. *Clin.Endo.* **1974**, 3, 69-96.
4. Avvakumov, GV. *J.Bio.Chem.* **2001**, 276, 34453-34457.
5. Lapidus, L. *Clin.Chem.* **1986**, 32, 146-152.
6. Tsai, EC. *Diabetes Care*, **2004**, 27, 861.
7. Thaler, M. *Clin.Chem.* **2005**, 51, 401-407.
8. Vermeulen, A. *J.Clin.Endo.* **1999**, 84, 3666.
9. Sodergard, R. *J.Steroid Biochem.* **1982**, 16, 801-810.
10. Mayo Clinic: Mayo Medical Laboratories Web Site.

< [Revision: 0 Date: 2013-OCT-04 DCO: na  
Product Code: 9125-300](http://www.mayomedicallaboratories.com/test-catalog/Clinical+and+Interpretive/9285></a>.</p>
</div>
<div data-bbox=)

For Orders and Inquiries, please contact



Tel: +1 949.951.2685 Email: [info@monobind.com](mailto:info@monobind.com)  
Fax: +1 949.951.3539 Web: [www.monobind.com](http://www.monobind.com)

Please visit our website to learn more about our other interesting products and services.



## Meridian Healthcare srl

Via Caronda, 446 SC/A - 95129 Catania - Italy  
Tel. +39 095 725 68 69 - Fax.: +39 095 725 44 54  
[info@meridianhealthcare.it](mailto:info@meridianhealthcare.it)  
[www.meridianhealthcare.it](http://www.meridianhealthcare.it)